

eHealth Solutions / Release VA36A / 2018-12-07 / Revision 83014

DICOM Conformance Statement

Distributed by:

SIEMENS

Contents

1	Introduction	1
1.1	Purpose	1
1.2	Intended Audience	1
1.3	Remarks	1
2	Implementation Model	2
2.1	Application Data Flow Diagram	2
2.1.1	XDS-I Source Adaptor Application Data Flow Diagram	2
2.1.2	Consumer App Data Flow Diagram	3
2.2	Functional Definition of AEs	3
2.2.1	XDS-I Source Adaptor	3
2.2.2	The Consumer App	3
3	Application Entity Specification	4
3.1	XDS-I Source Adaptor	4
3.1.1	Supported SOP Classes for XDS-I Source Adaptor AE	4
3.1.2	Association Establishment Policies	7
3.1.2.1	General	7
3.1.2.2	Number of Associations	7
3.1.2.3	Asynchronous Nature	7
3.2	Consumer App	7
3.2.1	Supported SOP Classes for Consumer App AE	7
4	Communication Profiles	9
4.1	Supported Communication Stacks	9
5	Security Profiles	10
5.1	Audit Trail Message Format Profile	10
6	Configuration	12
7	Supported Character Sets	13
8	Mapping from DICOM to XDS	14

1 Introduction

1.1 Purpose

This document is meant to formally describe the DICOM capabilities of this product. XDS-I Source Adaptor adds XDS-I.b capabilities to PACS systems. It provides DICOM interfaces to register images to a IHE XDS infrastructure and WebService interfaces to allow the retrieve of registered images with any XDS-I.b Consumer. The ConsumerApplication is a Web-Application to display DICOM images registered in a IHE XDS infrastructure. It also provides a functionality to store extramural DICOM studies to a local PACS system.

1.2 Intended Audience

This document is intended for network administrators, hospital staff or software designers who are responsible for system integration and/or software design. The terms used in this document are defined according to the DICOM- and XDS standard. It is thus assumed that the user is familiar with the terminology and concepts that are used in the DICOM 3.0 standard and the IHE ITI/RAD framework. If not, readers are asked to first read the appropriate parts of the DICOM standard and the IHE ITI/RAD framework themselves.

1.3 Remarks

The document should help to validate the integration of eHealth Solutions within a DICOM environment. It is not intended to replace the validation with other DICOM equipment to ensure proper exchange of information. The interoperability of the integration has to be secured separately. The user should be aware of the following important issues:

- The comparison of different Conformance Statements should be the first step towards the assessment of the interoperability within a DICOM environment
- Test procedures should be defined to validate the desired level of connectivity.


2 Implementation Model

2.1 Application Data Flow Diagram

2.1.1 XDS-I Source Adaptor Application Data Flow Diagram

The XDS-I Source Adaptor is implemented as a single AE which provides a set of different services. For each of these services the XDS-I Source Adaptor can act in a specific role.


Figure 1: XDS-I Source Adaptor Application Data Flow Diagram.



2.1.2 Consumer App Data Flow Diagram

Provides a web based user interface for displaying DICOM instances. It also provides the functionality to store images to a local PACS system.

Figure 2: Consumer App Data Flow Diagram.



2.2 Functional Definition of AEs

2.2.1 XDS-I Source Adaptor

The XDS-I Source Adaptor AE waits for another application to connect and then initiates a DICOM association. When another application connects, the XDS-I Source Adaptor AE expects it to be a DICOM application. XDS-I Source Adaptor AE implements several DICOM Service Classes. In total the following services are provided by this AE:

- Verification SCP answers communication tests from remote applications – C-ECHO
- Storage SCP implements the answer to external C-STORE requests. It is able to receive incoming DICOM image files sent by remote DICOM applications (e.g., PACS systems) and register them to a XDS infrastructure. Currently the XDS-I Source Adaptor only supports DICOM instances with ImplicitVRLittleEndian and ExplicitVRLittleEndian TransferSyntax.
- The Query/Retrieve SCP implements the answer to C-FIND, C-MOVE and C-GET requests. Remote applications can request queries on Patient-, Study-, Series- or Image-level using the Patient Root or Study Root query model. The XDS-I Source Adaptor forwards the request to the desired DICOM System. In case of a C-Move Request the XDS-I Source Adaptor replaces the move-destination with his own AET. The incoming C-Store then will be forwarded to the move-destination provided by the calling system. In this case the XDS-I Source Adaptor acts as a DICOM proxy.

2.2.2 The Consumer App

The **Consumer App** AE is able to render images available in a XDS infrastructure. Currently Structured Reports and Presentation States are not supported. It also implements the C-Store SCU to store DICOM instances to local PACS systems.

3 Application Entity Specification

3.1 XDS-I Source Adaptor

3.1.1 Supported SOP Classes for XDS-I Source Adaptor AE

Table 1: Supported Storage SOP classes.

SOP Class Name	SOP Class UID	SCP	SCU
Stored Print Storage	1.2.840.10008.5.1.1.1.27	Y	Y
Hardcopy Grayscale Image Storage	1.2.840.10008.5.1.1.1.29	Y	Y
Hardcopy Color Image Storage	1.2.840.10008.5.1.1.1.30	Y	Y
Computed Radiography Image Storage	1.2.840.10008.5.1.4.1.1.1	Y	Y
Digital X-Ray Image Storage – For Presentation	1.2.840.10008.5.1.4.1.1.1.1	Y	Y
Digital X-Ray Image Storage – For Processing	1.2.840.10008.5.1.4.1.1.1.1.1	Y	Y
Digital Mammography X-Ray Image Storage – For Presentation	1.2.840.10008.5.1.4.1.1.1.2	Y	Y
Digital Mammography X-Ray Image Storage – For Processing	1.2.840.10008.5.1.4.1.1.1.2.1	Y	Y
Digital Intra-oral X-Ray Image Storage – For Presentation	1.2.840.10008.5.1.4.1.1.1.3	Y	Y
Digital Intra-oral X-Ray Image Storage – For Processing	1.2.840.10008.5.1.4.1.1.1.3.1	Y	Y
CT Image Storage	1.2.840.10008.5.1.4.1.1.2	Y	Y
Enhanced CT Image Storage	1.2.840.10008.5.1.4.1.1.2.1	Y	Y
Legacy Converted Enhanced CT Image Storage	1.2.840.10008.5.1.4.1.1.2.2	Y	Y
Ultrasound Multiframe Image Storage (Retired)	1.2.840.10008.5.1.4.1.1.3	Y	Y
Ultrasound Multiframe Image Storage	1.2.840.10008.5.1.4.1.1.3.1	Y	Y
MR Image Storage	1.2.840.10008.5.1.4.1.1.4	Y	Y
Enhanced MR Image Storage	1.2.840.10008.5.1.4.1.1.4.1	Y	Y
MR Spectroscopy Storage	1.2.840.10008.5.1.4.1.1.4.2	Y	Y
Enhanced MR Color Image Storage SOP Class	1.2.840.10008.5.1.4.1.1.4.3	Y	Y
Legacy Converted Enhanced MR Image Storage	1.2.840.10008.5.1.4.1.1.4.4	Y	Y
Nuclear Medicine Image Storage (Retired)	1.2.840.10008.5.1.4.1.1.5	Y	Y
Ultrasound Image Storage (Retired)	1.2.840.10008.5.1.4.1.1.6	Y	Y
Ultrasound Image Storage	1.2.840.10008.5.1.4.1.1.6.1	Y	Y
Enhanced US Volume Storage	1.2.840.10008.5.1.4.1.1.6.2	Y	Y
Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7	Y	Y
Multiframe Single Bit Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.1	Y	Y
Multiframe Grayscale Byte Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.2	Y	Y
Multiframe Grayscale Word Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.3	Y	Y
Multiframe True Color Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.4	Y	Y
Standalone Overlay Storage	1.2.840.10008.5.1.4.1.1.8	Y	Y
Standalone Curve Storage	1.2.840.10008.5.1.4.1.1.9	Y	Y
Waveform Storage – Trial (Retired)	1.2.840.10008.5.1.4.1.1.9.1	Y	Y
12-lead ECG Waveform Storage	1.2.840.10008.5.1.4.1.1.9.1.1	Y	Y
General ECG Waveform Storage	1.2.840.10008.5.1.4.1.1.9.1.2	Y	Y
Ambulatory ECG Waveform Storage	1.2.840.10008.5.1.4.1.1.9.1.3	Y	Y
Hemodynamic Waveform Storage	1.2.840.10008.5.1.4.1.1.9.2.1	Y	Y
Cardiac Electrophysiology Waveform Storage	1.2.840.10008.5.1.4.1.1.9.3.1	Y	Y
Basic Voice Audio Waveform Storage	1.2.840.10008.5.1.4.1.1.9.4.1	Y	Y
General Audio Waveform Storage	1.2.840.10008.5.1.4.1.1.9.4.2	Y	Y

Table 1: Supported Storage SOP classes. 

Table 1: Supported Storage SOP classes. 

SOP Class Name	SOP Class UID	SCP	SCU
Arterial Pulse Waveform Storage	1.2.840.10008.5.1.4.1.1.9.5.1	Y	Y
Respiratory Waveform Storage	1.2.840.10008.5.1.4.1.1.9.6.1	Y	Y
Standalone Modality LUT Storage	1.2.840.10008.5.1.4.1.1.10	Y	Y
Standalone VOILUT Storage	1.2.840.10008.5.1.4.1.1.11	Y	Y
Grayscale Softcopy Presentation State Storage SOP Class	1.2.840.10008.5.1.4.1.1.11.1	Y	Y
Color Softcopy Presentation State Storage SOP Class	1.2.840.10008.5.1.4.1.1.11.2	Y	Y
Pseudo-Color Softcopy Presentation State Storage SOP Class	1.2.840.10008.5.1.4.1.1.11.3	Y	Y
Blending Softcopy Presentation State Storage SOP Class	1.2.840.10008.5.1.4.1.1.11.4	Y	Y
XA/XRF Grayscale Softcopy Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.5	Y	Y
X-Ray Angiographic Image Storage	1.2.840.10008.5.1.4.1.1.12.1	Y	Y
Enhanced XA Image Storage	1.2.840.10008.5.1.4.1.1.12.1.1	Y	Y
X-Ray Fluoroscopy Image Storage	1.2.840.10008.5.1.4.1.1.12.2	Y	Y
Enhanced XRF Image Storage	1.2.840.10008.5.1.4.1.1.12.2.1	Y	Y
X-Ray Angiographic Bi-Plane Image Storage (Retired)	1.2.840.10008.5.1.4.1.1.12.3	Y	Y
X-Ray 3D Angiographic Image Storage	1.2.840.10008.5.1.4.1.1.13.1.1	Y	Y
X-Ray 3D Craniofacial Image Storage	1.2.840.10008.5.1.4.1.1.13.1.2	Y	Y
Breast Tomosynthesis Image Storage	1.2.840.10008.5.1.4.1.1.13.1.3	Y	Y
Breast Projection X-Ray Image Storage For Presentation	1.2.840.10008.5.1.4.1.1.13.1.4	Y	Y
Breast Projection X-Ray Image Storage For Processing	1.2.840.10008.5.1.4.1.1.13.1.5	Y	Y
Intravascular Optical Coherence Tomography Image Storage – For Presentation	1.2.840.10008.5.1.4.1.1.14.1	Y	Y
Intravascular Optical Coherence Tomography Image Storage – For Processing	1.2.840.10008.5.1.4.1.1.14.2	Y	Y
Nuclear Medicine Image Storage	1.2.840.10008.5.1.4.1.1.20	Y	Y
Raw Data Storage	1.2.840.10008.5.1.4.1.1.66	Y	Y
Spatial Registration Storage	1.2.840.10008.5.1.4.1.1.66.1	Y	Y
Spatial Fiducials Storage	1.2.840.10008.5.1.4.1.1.66.2	Y	Y
Deformable Spatial Registration SOP Class	1.2.840.10008.5.1.4.1.1.66.3	Y	Y
Segmentation SOP Class	1.2.840.10008.5.1.4.1.1.66.4	Y	Y
Surface Segmentation Storage	1.2.840.10008.5.1.4.1.1.66.5	Y	Y
Real World Value Mapping Storage	1.2.840.10008.5.1.4.1.1.67	Y	Y
Surface Scan Mesh Storage	1.2.840.10008.5.1.4.1.1.68.1	Y	Y
Surface Scan Point Cloud Storage	1.2.840.10008.5.1.4.1.1.68.2	Y	Y
VL Image Storage (Retired)	1.2.840.10008.5.1.4.1.1.77.1	Y	Y
VL Endoscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.1	Y	Y
Video Endoscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.1.1	Y	Y
VL Microscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.2	Y	Y
Video Microscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.2.1	Y	Y
VL Slide Coordinates Microscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.3	Y	Y
VL Photographic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.4	Y	Y
Video Photographic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.4.1	Y	Y
Ophthalmic Photography 8 Bit Image Storage	1.2.840.10008.5.1.4.1.1.77.1.5.1	Y	Y
Ophthalmic Photography 16 Bit Image Storage	1.2.840.10008.5.1.4.1.1.77.1.5.2	Y	Y
Stereometric Relationship Storage	1.2.840.10008.5.1.4.1.1.77.1.5.3	Y	Y
Ophthalmic Tomography Image Storage	1.2.840.10008.5.1.4.1.1.77.1.5.4	Y	Y
VL Whole Slide Microscopy Image Storage	1.2.840.10008.5.1.4.1.1.77.1.6	Y	Y
VL Multi-frame Image Storage – Trial (Retired)	1.2.840.10008.5.1.4.1.1.77.2	Y	Y
Lensometry Measurements Storage	1.2.840.10008.5.1.4.1.1.78.1	Y	Y

Table 1: Supported Storage SOP classes. 

Table 1: Supported Storage SOP classes. 

SOP Class Name	SOP Class UID	SCP	SCU
Autorefracton Measurements Storage	1.2.840.10008.5.1.4.1.1.78.2	Y	Y
Keratometry Measurements Storage	1.2.840.10008.5.1.4.1.1.78.3	Y	Y
Subjective Refraction Measurements Storage	1.2.840.10008.5.1.4.1.1.78.4	Y	Y
Visual Acuity Measurements Storage	1.2.840.10008.5.1.4.1.1.78.5	Y	Y
Spectacle Prescription Report Storage	1.2.840.10008.5.1.4.1.1.78.6	Y	Y
Ophthalmic Axial Measurements Storage	1.2.840.10008.5.1.4.1.1.78.7	Y	Y
Intraocular Lens Calculations Storage	1.2.840.10008.5.1.4.1.1.78.8	Y	Y
Macular Grid Thickness and Volume Report	1.2.840.10008.5.1.4.1.1.79.1	Y	Y
Ophthalmic Visual Field Static Perimetry Measurements Storage	1.2.840.10008.5.1.4.1.1.80.1	Y	Y
Ophthalmic Thickness Map Storage	1.2.840.10008.5.1.4.1.1.81.1	Y	Y
Corneal Topography Map Storage	1.2.840.10008.5.1.4.1.1.82.1	Y	Y
Text SR Storage – Trial (Retired)	1.2.840.10008.5.1.4.1.1.88.1	Y	Y
Audio SR Storage – Trial (Retired)	1.2.840.10008.5.1.4.1.1.88.2	Y	Y
Detail SR Storage – Trial (Retired)	1.2.840.10008.5.1.4.1.1.88.3	Y	Y
Comprehensive SR Storage – Trial (Retired)	1.2.840.10008.5.1.4.1.1.88.4	Y	Y
Basic Text SR Storage	1.2.840.10008.5.1.4.1.1.88.11	Y	Y
Enhanced SR Storage	1.2.840.10008.5.1.4.1.1.88.22	Y	Y
Comprehensive SR Storage	1.2.840.10008.5.1.4.1.1.88.33	Y	Y
Comprehensive 3D SR Storage	1.2.840.10008.5.1.4.1.1.88.34	Y	Y
Procedure Log Storage	1.2.840.10008.5.1.4.1.1.88.40	Y	Y
Mammography CAD SR Storage	1.2.840.10008.5.1.4.1.1.88.50	Y	Y
Key Object Selection Document Storage	1.2.840.10008.5.1.4.1.1.88.59	Y	Y
Chest CAD SR Storage	1.2.840.10008.5.1.4.1.1.88.65	Y	Y
X-Ray Radiation Dose SR Storage	1.2.840.10008.5.1.4.1.1.88.67	Y	Y
Radiopharmaceutical Radiation Dose SR Storage	1.2.840.10008.5.1.4.1.1.88.68	Y	Y
Colon CAD SR	1.2.840.10008.5.1.4.1.1.88.69	Y	Y
Implantation Plan SR Document Storage	1.2.840.10008.5.1.4.1.1.88.70	Y	Y
Encapsulated PDF Storage	1.2.840.10008.5.1.4.1.1.104.1	Y	Y
Encapsulated CDA Storage	1.2.840.10008.5.1.4.1.1.104.2	Y	Y
PET Image Storage	1.2.840.10008.5.1.4.1.1.128	Y	Y
Legacy Converted Enhanced PET Image Storage	1.2.840.10008.5.1.4.1.1.128.1	Y	Y
PET Curve Storage	1.2.840.10008.5.1.4.1.1.129	Y	Y
Enhanced PET Image Storage	1.2.840.10008.5.1.4.1.1.130	Y	Y
Basic Structured Display Storage	1.2.840.10008.5.1.4.1.1.131	Y	Y
RT Image Storage	1.2.840.10008.5.1.4.1.1.481.1	Y	Y
RT Dose Storage	1.2.840.10008.5.1.4.1.1.481.2	Y	Y
RT Structure Set Storage	1.2.840.10008.5.1.4.1.1.481.3	Y	Y
RT Beams Treatment Record Storage	1.2.840.10008.5.1.4.1.1.481.4	Y	Y
RT Plan Storage	1.2.840.10008.5.1.4.1.1.481.5	Y	Y
RT Brachy Treatment Record Storage	1.2.840.10008.5.1.4.1.1.481.6	Y	Y
RT Treatment Summary Record Storage	1.2.840.10008.5.1.4.1.1.481.7	Y	Y
RT Ion Plan Storage	1.2.840.10008.5.1.4.1.1.481.8	Y	Y
RT Ion Beams Treatment Record Storage	1.2.840.10008.5.1.4.1.1.481.9	Y	Y
DICOS CT Image Storage	1.2.840.10008.5.1.4.1.1.501.1	Y	Y
DICOS Digital X-Ray Image Storage – For Presentation	1.2.840.10008.5.1.4.1.1.501.2.1	Y	Y
DICOS Digital X-Ray Image Storage – For Processing	1.2.840.10008.5.1.4.1.1.501.2.2	Y	Y
DICOS Threat Detection Report Storage	1.2.840.10008.5.1.4.1.1.501.3	Y	Y
Eddy Current Image Storage	1.2.840.10008.5.1.4.1.1.601.1	Y	Y

Table 1: Supported Storage SOP classes. 

Table 1: Supported Storage SOP classes. 

SOP Class Name	SOP Class UID	SCP	SCU
Eddy Current Multi-frame Image Storage	1.2.840.10008.5.1.4.1.1.601.2	Y	Y
RT Beams Delivery Instruction Storage – Trial (Retired)	1.2.840.10008.5.1.4.34.1	Y	Y
RT Beams Delivery Instruction Storage	1.2.840.10008.5.1.4.34.7	Y	Y
Generic Implant Template Storage	1.2.840.10008.5.1.4.43.1	Y	Y
Implant Assembly Template Storage	1.2.840.10008.5.1.4.44.1	Y	Y
Implant Template Group Storage	1.2.840.10008.5.1.4.45.1	Y	Y

3.1.2 Association Establishment Policies

3.1.2.1 General

The XDS-I Source Adaptor AE supports plain TCP and TLS encrypted communication. For each kind of transport the server provides an arbitrary number of listen ports. All these ports are equivalent and provide the same services. The maximum PDU size accepted is 16384.

3.1.2.2 Number of Associations

The XDS-I Source Adaptor has a configured association limit on each listen port.

3.1.2.3 Asynchronous Nature

The XDS-I Source Adaptor does not support asynchronous communication. Multiple outstanding transactions are not supported. It allows up to one invoked and one performed operation on an Association (it is synchronous). An asynchronous mode of operation is not supported.

Table 2: Implementation Classes

Class	Value
Implementation Class UID	1.2.40.0.13.1.1
Implementation Version Name	dcm4che-2.0

3.2 Consumer App

3.2.1 Supported SOP Classes for Consumer App AE

Table 3: Supported SOP Classes for the Consumer App AE.

SOP Class Name	SOP Class UID	SCP	SCU
Computed Radiography Image Storage	1.2.840.10008.5.1.4.1.1.1	Y	Y
Digital X-Ray Image Storage For Presentation	1.2.840.10008.5.1.4.1.1.1.1	Y	Y
Digital X-Ray Image Storage For Processing	1.2.840.10008.5.1.4.1.1.1.1.1	Y	Y
Digital Mammography X-Ray Image Storage For Presentation	1.2.840.10008.5.1.4.1.1.1.2	Y	Y

Table 3: Supported SOP Classes for the Consumer App AE. 

Table 3: Supported SOP Classes for the Consumer App AE. 

SOP Class Name	SOP Class UID	SCP	SCU
Digital Mammography X-Ray Image Storage For Processing	1.2.840.10008.5.1.4.1.1.1.2.1	Y	Y
Digital Intraoral X-Ray Image Storage For Presentation	1.2.840.10008.5.1.4.1.1.1.3	Y	Y
Digital Intraoral X-Ray Image Storage For Processing	1.2.840.10008.5.1.4.1.1.1.3.1	Y	Y
CT Image Storage	1.2.840.10008.5.1.4.1.1.2	Y	Y
Enhanced CT Image Storage	1.2.840.10008.5.1.4.1.1.2.1	Y	Y
Ultrasound Multiframe Image Storage	1.2.840.10008.5.1.4.1.1.3.1	Y	Y
MR Image Storage	1.2.840.10008.5.1.4.1.1.4	Y	Y
Enhanced MR Image Storage	1.2.840.10008.5.1.4.1.1.4.1	Y	Y
Enhanced MR Color Image Storage	1.2.840.10008.5.1.4.1.1.4.3	Y	Y
Ultrasound Image Storage	1.2.840.10008.5.1.4.1.1.6.1	Y	Y
Enhanced US Volume Storage	1.2.840.10008.5.1.4.1.1.6.2	Y	Y
Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7	Y	Y
Multiframe Single Bit Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.1	Y	Y
Multiframe Grayscale Byte Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.2	Y	Y
Multiframe Grayscale Word Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.3	Y	Y
Multiframe True Color Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.4	Y	Y
X-Ray Angiographic Image Storage	1.2.840.10008.5.1.4.1.1.12.1	Y	Y
X-Ray Radiofluoroscopic Image Storage	1.2.840.10008.5.1.4.1.1.12.2	Y	Y
X-Ray 3D Angiographic Image Storage	1.2.840.10008.5.1.4.1.1.13.1.1	Y	Y
X-Ray 3D Craniofacial Image Storage	1.2.840.10008.5.1.4.1.1.13.1.2	Y	Y
Breast Tomosynthesis Image Storage	1.2.840.10008.5.1.4.1.1.13.1.3	Y	Y
Nuclear Medicine Image Storage	1.2.840.10008.5.1.4.1.1.20	Y	Y
VL Endoscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.1	Y	Y
VL Microscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.2	Y	Y
VL Slide Coordinates Microscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.3	Y	Y
VL Photographic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.4	Y	Y
Ophthalmic Photography 8Bit Image Storage	1.2.840.10008.5.1.4.1.1.77.1.5.1	Y	Y
Ophthalmic Photography 16Bit Image Storage	1.2.840.10008.5.1.4.1.1.77.1.5.2	Y	Y
Ophthalmic Tomography Image Storage	1.2.840.10008.5.1.4.1.1.77.1.5.4	Y	Y
Positron Emission Tomography Image Storage	1.2.840.10008.5.1.4.1.1.128	Y	Y
Enhanced PET Image Storage	1.2.840.10008.5.1.4.1.1.130	Y	Y
RT Image Storage	1.2.840.10008.5.1.4.1.1.481.1	Y	Y
Hardcopy Grayscale Image Storage SOP Class Retired	1.2.840.10008.5.1.1.29	Y	Y
Hardcopy Color Image Storage SOP Class Retired	1.2.840.10008.5.1.1.30	Y	Y
Nuclear Medicine Image Storage Retired	1.2.840.10008.5.1.4.1.1.5	Y	Y
Ultrasound Multiframe Image Storage Retired	1.2.840.10008.5.1.4.1.1.3	Y	Y
Ultrasound Image Storage Retired	1.2.840.10008.5.1.4.1.1.6	Y	Y
X-Ray Angiographic BiPlane Image Storage Retired	1.2.840.10008.5.1.4.1.1.12.3	Y	Y
VL Image Storage Trial Retired	1.2.840.10008.5.1.4.1.1.77.1	Y	Y
VL Multiframe Image Storage Trial Retired	1.2.840.10008.5.1.4.1.1.77.2	Y	Y

4 Communication Profiles

4.1 Supported Communication Stacks

XDS-I Source Adaptor and **Consumer App** provide plain TCP (see DICOM 2011 PS 3.8, 9) and TLS encrypted communication (see DICOM 2011 PS 3.15, B.1). They uses dcm4che for their communications.

5 Security Profiles

5.1 Audit Trail Message Format Profile

To help assure healthcare privacy and security in automated systems, information about the data usage need to be collected. This data will be reviewed by administrative staff to verify that healthcare data is being used in accordance with the healthcare provider's data security requirements and to establish accountability for data use. This data collection and review process is called security auditing and the data itself comprises the audit trail. Audit trails can be used for surveillance purposes to detect when interesting events might be happening that warrant further investigation.

Auditing in eHealth Solutions is implemented according to the IHE profile Audit Trail and Node Authentication (part Audit Trail) which is based on RFC 3881 (audit message XML format) and DICOM 2011 PS 3.15, A.5. Auditing is restricted to events regarding patients, documents, user authentications and audit configuration modifications:

- search for patients or documents
- creation and modification of patients
- import, modification and export of documents

The processing of audit messages works asynchronously — events are recorded immediately, however, the resulting audit messages are queued and sent to an Audit Record Repository via TCP or WebServices in periodic time intervals (the frequency of the intervals is configurable). If required, audit messages can be analyzed with an Audit Record Viewer. To use audit messages for effective system analyses, it is necessary that each audit message can be uniquely associated with a certain event. To this end each audit message provides various information. The most important values are:

- event ID,
- date and time of the event,
- status of the event,
- user IDs,
- application IDs,
- object IDs,
- audit trail ID (used to aggregate audit messages in order to reconstruct audit trails), and
- audit source ID.

Within eHealth Solutions, the events recorded by audit messages are primarily identified via event IDs. However, since most event IDs represent a group of events rather than a single event, it is often necessary to explore audit messages in detail to identify the reported events. In the following, the event IDs used by eHealth Solutions as well as some short instructions on how to identify events are listed below:

110106 (Export):

This event ID indicates that a DICOM study was registered to the XDS infrastructure.

■ 110104 (DICOM Instances Transferred):

The event ID 110104 is used to state the end of a transfer of DICOM images from the XDS-I Source to a XDS-I consumer.

■ 110103 (DICOM Instances Accessed):

This event ID is used in audit messages which are generated if DICOM images are received by the Consumer.

To support the analysis of audit messages within eHealth Solutions, audit messages are organized in so called audit trails. An audit trail represents a group of audit messages which have been created during the execution of a certain action (e.g. registering a DICOM study). Since it might happen that several sub-actions have to be performed in order to complete an action within an audit trail, audit messages are classified into sub and main audit messages. Thereby, sub audit messages are used to audit necessary sub-actions whereas main audit messages are used to audit sufficient sub-actions. To distinguish between different audit trails, each audit trail is equipped with a unique audit trail ID.

6 Configuration

The eHealth Solutions XDS-I SourceAdaptor and the **Consumer App** uses a global XML based configuration. The configuration is provided by the ConfigurationRepository which is queried periodically by the services. If the ConfigurationRepository is unavailable a local cached configuration is used.

7 Supported Character Sets

All character sets from DICOM 2011 PS 3.3, C.12.1.1.2 are supported with and without code extension techniques.

8 Mapping from DICOM to XDS

For proper processing of DICOM studies in eHealth Solutions, it is important to map the DICOM metadata in a corresponding way. This section provides an overview of correlations between DICOM tags and XDS fields.

Table 4: Mapping of Document Metadata

DICOM Tag	DICOM Name	VR	XDS	O/R	Adjustable
0008,1030	StudyDescription	LO	Title	R	true
0032,4000	StudyComments	LT	Comments	O	true
0008,0012	InstanceCreationDate	DA	CreationTime	R ¹	true
0008,0013	InstanceCreationTime	TM			
0008,0020	StudyDate	DA	ServiceStartTime	R	true
0008,0030	StudyTime	TM		O ²	
0008,2218	AnatomicRegionSequence	SQ	EventCodes	O	false
> 0008,0100	CodeValue	SH	Value	R ³	false
> 0008,0102	CodingSchemeDesignator	SH	CodingScheme	R ³	false
> 0008,0104	CodeMeaning	LO	DisplayName	R ³	false
0008,1050	PerformingPhysicianName	PN ⁴	Author	R	true
0008,0080	InstitutionName	LO	Institution.Name	R	true

Table 5: Mapping of Patient Metadata

DICOM Tag	DICOM Name	VR	XDS	O/R	Adjustable
0010,0010	PatientName	PN ⁴		R	false
0010,0020	PatientID	LO	SourcePatientID ⁵	R	false
0010,0030	PatientBirthDate	DA	Birthdate	R	false
0010,0040	PatientSex	CS	Gender	R	false

The remaining metadata (ClassCodes, PracticeSettingCodes, Confidentiality Codes, FormatCodes, ContentTypeCodes) have to be mapped in the XDI - Source configuration. In addition, eHealth Solutions offers the possibility to overwrite DICOM tags for certain AETs (Called as well as Calling) by default. This allows for a better handling of shortcomings, for instance, when source systems do not support the usage of certain tags.

¹Can be defaulted to current time.

²When missing, the value is set to 00:00:00.

³Only required if AnatomicRegionSequence is set in DICOM.

⁴At least the family name and given name are required. See DICOM PS 3.5, table 6.2-1 for details.

⁵The OID for the SourcePatientID will be mapped from InstitutionName.

Distributed by

Siemens Healthcare GmbH
Henkestr. 127
91052 Erlangen
Germany
Phone: +49 9131 84-0
siemens.com/healthcare

Legal Manufacturer

ITH icoserve technology for healthcare GmbH
Innrain 98
6020 Innsbruck
Austria
Phone: +43 512 89059